


# ASSET™ EZ4 Dry Samplers for Isocyanates

First Class Sampling and Analysis



The life science business of Merck KGaA Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

**Supelco®**  
Analytical Products

## Two ASSET™ EZ4 Samplers Available

ASSET™ EZ4 dry samplers for isocyanates are easy-to-use dry samplers offering exceptional sensitivity for collection and measurement of vapor phase and particulate isocyanates. The ASSET™ EZ4-NCO dry sampler collects the full range of isocyanate monomers and oligomers; while the ASSET™ EZ-ICA sampler is designed to collect isocyanic acid (ICA) and methyl isocyanate (MIC) at low levels.

The science behind the ASSET™ EZ4 sampler is based on dibutylamine (DBA) impregnated glass fiber filters housed in a denuder and filter cassette. This ensures both the vapor phase and aerosol/particulate isocyanates are completely derivatized and form stable derivatives. The sampling method is fully validated according to ISO 17734-1:2013.

### Key Features and Benefits

- Easy to use and safe to wear
- Increases workplace productivity during industrial hygiene (IH) sampling, as work shifts are not interrupted for sampler exchange
- Reduces the cost of analysis through a reduction in required samplers to cover a work shift
- Ability to achieve reliable, low detection limits of 50-100x below existing methods
- Only device to measure vapor phase and particulate isocyanates
- Fully supported by certified reference materials for LC-MS analysis (calibration and internal standards)
- No interferences, no field extraction
- Reduces shipping and handling charges
- 2-year shelf life



ASSET™ EZ4-NCO Dry Sampler



ASSET™ EZ4-ICA Dry Sampler

	ASSET™ EZ4-NCO Dry Sampler	ASSET™ EZ4-ICA Dry Sampler
Sampling Flow Rate	From 20 mL/min - 850 mL/min	From 20 mL/min - 200 mL/min
Sampling Time	From 5 minutes up to 12 hrs.	From 5 minutes up to 4 hrs.
Storage	Ambient (before and after sampling)	Refrigeration (before sampling); Ambient (after sampling)
Stability	4 weeks after sampling	2 weeks after sampling


### Did You Know?

- In 2013, several piperazine compounds, a common chemical used in the derivatization process of isocyanates in several existing sampling devices, were added to the banned substance list in the United Kingdom.
- Several common isocyanate sampling devices require the use of toluene, a Class 3 hazardous chemical, as a component of sampling, requiring the high cost of compliance shipping hazardous chemicals to the laboratory.

### Chromatogram of DBA Standard

(equivalent 0.20 µg/mL each underivatized monomer)

column: Ascentis® Express C18, 15 cm x 4.6 mm I.D., 2.7 µm particles (53829-U)  
mobile phase: (A) 5:95 acetonitrile:water w/ 0.05% formic acid; (B) 95:5 acetonitrile:water w/ 0.05% formic acid  
gradient: 40% to 70% B in 3 min; 70% to 90% B in 2 min; hold at 90% for 6 min, 90% to 40% in 0.1 min, hold at 40% for 3.9 min  
flow rate: 1 mL/min  
column temp.: 35 °C  
detector: AB3200 QTrap™ MS, ESI(+), SIM  
injection: 2 µL



Description	Qty.	Cat. No.
<b>Sampling Device</b>		
ASSET™ EZ4-NCO Dry Sampler	10	5027-U
	50	5028-U
Sample kit – Includes 2 samplers and fittings		
	2	5047-U
ASSET™ EZ4-ICA Dry Sampler	10	28517-U
	40	28518-U
Calibration Fitting		
	10	5048-U
<b>HPLC Columns</b>		
LC-MS-MS Ascentis® Express C18, 5 cm x 2.1 mm I.D., 2.7 µm	1	53822-U
LC-MS Ascentis® Express C18, 15 cm x 4.6 mm I.D., 2.7 µm	1	53829-U

## Certified Reference Materials

Our isocyanate-dibutylamine (DBA) calibration and isocyanate-dibutylamine-d<sub>9</sub> internal standard mixes were developed for use with the Supelco® ASSET EZ4 dry samplers. These certified reference materials (CRMs) are produced in an ISO/IEC 17025 and ISO 17034 accredited lab and are traceable to a NIST Standard Reference Material (SRM). A certificate of analysis fulfilling the requirements of ISO Guide 31 is provided with each CRM.

Description	Concentration	Synonym	Pkg.	Cat. No.
<b>Mixtures</b>				
TraceCERT® DBA Isocyanate Monomers Mix	in acetone:methanol (99:1) (varied conc.)* Isocyanic acid-di-n-butylamine Ethyl isocyanate-di-n-butylamine derivative 1,6-Hexamethylene diisocyanate-2(di-n-butylamine) derivative Isophorone isocyanate-2(di-n-butyl amine) Isomer 1 derivative Isophorone isocyanate-2(di-n-butyl amine) Isomer 2 derivative 4,4'-Methylenediphenyl diisocyanate-2(di-n-butylamine) derivative Methyl isocyanate-di-n-butylamine derivative Phenyl isocyanate-di-n-butylamine derivative Propyl isocyanate-di-n-butylamine derivative 2,4-Toluene diisocyanate-2(di-n-butylamine) derivative 2,6-Toluene diisocyanate-2(di-n-butylamine) derivative	ICA-DBA EIC-DBA HDI-2(DBA) IPDI-2(DBA) IPDI-2(DBA) 4,4'-MDI-2(DBA) MIC-DBA PhI-DBA PIC-DBA 2,4-TDI-2(DBA) 2,6-TDI-2(DBA)	1 x 1 mL	CRM40569
TraceCERT® d <sub>9</sub> -DBA Isocyanate Monomers Mix	in acetone (varied conc.)* Isocyanic acid-di-n-butylamine-d <sub>9</sub> derivative Ethyl isocyanate-di-n-butylamine-d <sub>9</sub> derivative 1,6-Hexamethylene diisocyanate-2(di-n-butylamine-d <sub>9</sub> ) derivative Isophorone diisocyanate-2(di-n-butylamine-d <sub>9</sub> ) Isomer 1 derivative Isophorone diisocyanate-2(di-n-butylamine-d <sub>9</sub> ) Isomer 2 derivative 4,4'-Methylenediphenyl diisocyanate-2(di-n-butylamine-d <sub>9</sub> ) derivative Methyl isocyanate-di-n-butylamine-d <sub>9</sub> derivative Phenyl isocyanate-di-n-butylamine-d <sub>9</sub> derivative Propyl isocyanate-di-n-butylamine-d <sub>9</sub> derivative 2,4-Toluene diisocyanate-2(di-n-butylamine-d <sub>9</sub> ) derivative 2,6-Toluene diisocyanate-2(di-n-butylamine-d <sub>9</sub> ) derivative	ICA-DBA-d <sub>9</sub> EIC-DBA-d <sub>9</sub> HDI-2(DBA-d <sub>9</sub> ) IPDI-2(DBA-d <sub>9</sub> ) IPDI-2(DBA-d <sub>9</sub> ) 4,4'-MDI-2(DBA-d <sub>9</sub> ) MIC-DBA-d <sub>9</sub> PhI-DBA-d <sub>9</sub> PIC-DBA-d <sub>9</sub> 2,4-TDI-2(DBA-d <sub>9</sub> ) 2,6-TDI-2(DBA-d <sub>9</sub> )	1 x 1 mL	CRM40570
TraceCERT® HDI-DBA Oligomers Standard	in methanol (varied conc.) HDI-Monomer di-n-butylamine HDI-Uretidine di-n-butylamine HDI-Biuret di-n-butylamine HDI-Isocyanurate di-n-butylamine		1 x 1 mL	CRM40589
TraceCERT® HDI-DBA-d <sub>9</sub> Oligomers Standard	in methanol (varied conc.) HDI-Monomer di-n-butylamine-d <sub>9</sub> HDI-Uretidine di-n-butylamine-d <sub>9</sub> HDI-Biuret di-n-butylamine-d <sub>9</sub> HDI-Isocyanurate di-n-butylamine-d <sub>9</sub>		1 x 1 mL	CRM40590
TraceCERT® MDI-DBA Oligomers Standard	in methanol (varied conc.) Total MDI-di-n-butylamine monomers MDI-3-Ring-di-n-butylamine MDI-4-Ring-di-n-butylamine		1 x 1 mL	CRM40603
TraceCERT® MDI-DBA-d <sub>9</sub> Oligomers Standard	in methanol (varied conc.) MDI-Monomers-di-n-butylamine-d <sub>9</sub> MDI-3-Ring-di-n-butylamine-d <sub>9</sub> MDI-4-Ring-di-n-butylamine-d <sub>9</sub>		1 x 1 mL	CRM40604
TraceCERT® IPDI-DBA Oligomers Standard	in methanol (varied conc.) IPDI-Monomer-di-n-butylamine (Isomer 1) IPDI-Monomer-di-n-butylamine (Isomer 2) IPDI-Isocyanurate-di-n-butylamines Oligomers		1 x 1 mL	CRM40605
TraceCERT® IPDI-DBA-d <sub>9</sub> Oligomers Standard	in methanol (varied conc.) IPDI-Monomer-di-n-butylamine-d <sub>9</sub> (Isomer 1) IPDI-Monomer-di-n-butylamine-d <sub>9</sub> (Isomer 2) IPDI-Isocyanurate-di-n-butylamine-d <sub>9</sub> (Oligomers)		1 x 1 mL	CRM40606

\*Target is 100 µg/mL for individual components, with the exception of the mixes having isomer 1 and isomer 2. For isomers, the sum of the isomers should be 100 µg/mL. Factual concentration will be given in the lot specific certificate of analysis.

## References

- ISO 17734-1:2013, Determination of organonitrogen compounds in air using liquid chromatography and mass spectrometry — Part 1: Isocyanates using dibutylamine derivatives.
- ISO 17734-2:2013, Determination of organonitrogen compounds in air using liquid chromatography and mass spectrometry — Part 2: Amines and aminoisocyanates using dibutylamine and ethyl chloroformate derivatives.
- ISO/TR 17737:2007, Workplace air — Guidelines for selecting analytical methods for sampling and analyzing isocyanates in air.
- Brady J. M., et al. *Environ. Sci. Technol.* 2014, 48, 11405-11412
- Karlsson, D., Dalene, M., Scarping, G., Marand, A. *J. Environ. Monit.* 2001, 3, 432-436.
- HSE, WATCH/2008/4, ANNEX 2 "Assessment of the Potential for Isocyanic Acid and Other Monoisocyanates to Cause Respiratory Irritation and Sensitization".

# Supelco®

Analytical Products

Merck KGaA  
Frankfurter Strasse 250  
64293 Darmstadt, Germany



For more information, visit  
**SigmaAldrich.com/asset**

## To place an order or receive technical assistance

Order/Customer Service: [SigmaAldrich.com/order](https://SigmaAldrich.com/order)

Technical Service: [SigmaAldrich.com/techservice](https://SigmaAldrich.com/techservice)

Safety-related Information: [SigmaAldrich.com/safetycenter](https://SigmaAldrich.com/safetycenter)

**SigmaAldrich.com**

© 2020 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved. Merck, the vibrant M, Sigma-Aldrich, Supelco, Ascentis and ASSET are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.

Lit. No. MK\_FL5337EN Ver. 1.0 30191 02/2020