



QUALITY CONTROL RANGES

MILLIPLEX[®] MAP Human Complement Magnetic Bead Panel 2
Kit Catalog# HCMP2MAG-19K
Control Catalog # HCMP2-6019-2, Lot# HCM2-106 and HCM2-206

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
Complement C1q	Control 1	0.59 - 1.2	ng/mL
	Control 2	2.2 - 4.5	ng/mL
Complement C3	Control 1	2.0 - 4.1	ng/mL
	Control 2	7.7 - 16	ng/mL
Complement C3b/iC3b	Control 1	95 - 228	ng/mL
	Control 2	263 - 546	ng/mL
Complement C4	Control 1	3.9 - 8.2	ng/mL
	Control 2	18 - 38	ng/mL
Complement Factor B	Control 1	0.58 - 1.2	ng/mL
	Control 2	2.3 - 4.9	ng/mL
Complement Factor H	Control 1	3.2 - 6.7	ng/mL
	Control 2	12 - 26	ng/mL

HCMP2-6019-2: HCM2-106.206 (28 May 2021)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.