

Technical Data Sheet

Sabouraud Dextrose Agar + LTHTh - LI

Ordering number: 1.46052.0020 / 1.46052.0120

Sabouraud Dextrose Agar + LTHTh - LI is designed for the determination of the total aerobic count of yeasts and molds in air via active or passive air monitoring as well as fingerprints of personnel in controlled environments in the presence of disinfectant residues.

Ten settle plates each with a diameter of 90 mm are single-bagged in transparent, hydrogen peroxide impermeable sleeves (non-irradiated). The sleeves consist of polypropylene with a barrier of PE-EVOH-PE.

The formulation of the basic medium (Sabouraud Dextrose Agar) is prepared according to the recommendations of the current European, Japanese and United States Pharmacopoeia (EP, 2.6.12.; JP, 4.05 and USP, 61) and supplemented with neutralizers.

Mode of Action

Sabouraud Dextrose Agar (SDA) is a complex medium for cultivation and isolation of yeasts and molds as well as the absence test for *Candida albicans*. According to pharmacopoeia and ISO 18415, the neutralizers lecithin, polysorbate (Tween®) 80, histidine and sodium thiosulfate are suitable for neutralization of disinfectant residues containing the following active agents:

- Aldehydes
- Bis-biguanides
- Oxidizing compounds
- Parabens
- Phenolic compounds
- Quaternary ammonium compounds

The high concentration of dextrose in addition with the low pH promotes the growth, the formation of spores (conidia and sporangia) as well as the formation of pigments of yeasts and molds. On the other side the growth of bacteria is inhibited.

Typical Composition

Casein Peptone	5 g/l
Meat Peptone	5 g/l
Dextrose	40 g/l
Lecithin	0.7 g/l
Polysorbate (Tween®) 80	5 ml /l
Histidine	0.5 g/l
Sodium Thiosulfate	0.3 g/l
Agar	18 g/l

The appearance of the medium is clear and yellowish. The pH value is in the range of 5.4-5.8. The medium can be adjusted and/or supplemented according to the performance criteria required.

Application and Interpretation

Each plate is provided with a label including a data matrix code for paperless plate identification. The code consists of a two-dimensional 20-digit serial number, which harbors the following information:

Digits 1-3: here code 782 (corresponds to article 146052); digits 4-9: lot number; digits 10-14: batch specific individual number; digits 15-20: expiry date (YY/MM/DD).

Please check each agar plate before using it on sterility and pay attention to aseptic handling to avoid false positive results.

The plates may be used for passive or active air monitoring as described in USP chapter <1116> or ISO 14698. For active air sampling please follow the guidance of the air sampler. Typically, 1000 liter of air are sampled for quantification of CFU. The exposure time of opened settle plates should be validated with respect to the environmental conditions of the sampling area such as flow rates, temperatures and relative humidity to preclude desiccation. Afterwards the plates are closed and transferred to an incubator. To protect the plates from secondary contamination during transport and incubation outside of the Clean Room zone, sterile transport bags (article number 146509) may be used.

Several recommendations are given by different guidelines for incubation: according to USP <1116> the plates used for environmental monitoring should be incubated between 20 and 35 °C for not less than 72 hours. According to the FDA Aseptic Guide the plates for determination of the total yeast and mold count should be incubated at 20 to 25 °C for 5 to 7 days. Individual incubation conditions can be chosen and should be validated at the application side.

Finally, the number of CFU per plate is examined.

Grown colonies are recommended to be identified.

Storage and Shelf Life



The product can be used for sampling until the expiry date if stored upright, protected from light and properly sealed at +15 °C to +25 °C.

Condensation can be prevented by avoiding quick temperature shifts and mechanical stress.

The testing procedures as described on the CoA can be started up to the expiry date printed on the label.

Disposal

Please mind the respective regulations for the disposal of used culture medium (e.g. autoclave for 20 min at 121 °C, disinfect, incinerate etc.).

Quality Control

Control Strains	ATCC #	Inoculum CFU	Incubation	Expected Result Recovery in %
<i>Candida albicans</i>	10231	10-100	44-48 h at 20-25 °C	50-200
<i>Aspergillus brasiliensis</i>	16404	10-100	70-74 h at 20-25 °C	50-200

Please refer to the actual batch related Certificate of Analysis.

Literature

EU GMP Medicinal Products for Human and Veterinary use (2008): Annex1 Manufacture of Sterile Medicinal Products.

European Pharmacopoeia 9.0 (2016): 2.6.12. Microbial examination of non-sterile products (total viable aerobic count).

Guidance for Industry (2004): Sterile Drug Products Produced by Aseptic Processing - Current Good Manufacturing Practice.

ISO 14698-1:2003: Clean Rooms and associated controlled environments - Biocontamination control - Part 1: General principles and methods.

ISO 18415 (2017 [E]): Cosmetics – Microbiology – Detection of specified and non-specified microorganisms

Japanese Pharmacopoeia 16th edition (2011): 4.05 Microbial Limit Test.

PDA Technical Report No. 13 (2014 Revised): Fundamentals of an Environmental Monitoring Program.
United States Pharmacopoeia 341 NF 36 (2018): <61> Microbiological Examination of Non-Sterile Products: Microbial Enumeration Tests; <1116> Microbiological Control and Monitoring of Aseptic Processing Environments.

Ordering Information

Product	Cat. No.	Pack size
Sabouraud Dextrose Agar + LTHTh - LI	1.46052.0020	20 x 90 mm plates
Sabouraud Dextrose Agar + LTHTh - LI	1.46052.0120	120 x 90 mm plates

Merck KGaA
64271 Darmstadt, Germany
Fax: +49 (0) 61 51 / 72-60 80
mibio@merckgroup.com

Find contact information for your country
at: www.merckmillipore.com/offices

For Technical Service, please visit:
www.merckmillipore.com/techservice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and liability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any right of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

Merck, Millipore, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. Detailed information on trademarks is available via publicly accessible resources.
© 2019 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck operates as
MilliporeSigma in the U.S. and Canada.

Lit. No. TN5766EN00

Page 4 of 4

The Merck logo, consisting of the word "MERCK" in a bold, blue, sans-serif font.