

## Technical Data Sheet

# Sabouraud Dextrose Agar + LTHTh - ICR selective

Ordering number: 1.46016.0120

Sabouraud Dextrose Agar + LTHTh - ICR selective is designed for the determination of the total aerobic count of yeasts and molds in air via active or passive air monitoring as well as fingerprints of personnel in **Isolators** and **Clean Rooms**.

Ten settle plates each with a diameter of 90 mm are triple-bagged in transparent, hydrogen peroxide impermeable sleeves. The product is gamma-irradiated in the final packaging at a dose of 9-20 kGy. The sleeves consist of polypropylene with a barrier of PE-EVOH-PE.

To differentiate Tryptic soy agar (TSA) from Sabouraud Dextrose agar (SDA), SDA media are filled in pink colored dish (except ICR plus settle plates).

The formulation of the basic medium (Peptone Dextrose Agar) is prepared according to the recommendations of the current European, Japanese and United States Pharmacopoeia (EP, 2.6.12.; JP, 4.05 and USP, 61) and supplemented with neutralizers.

Further plate designs are available with the same media formulation:

- SDA Contact + LTHTh - ICR+ selective (article number 146538): 55 mm contact plates (pink dishes), triple-bagged, gamma-irradiated; intended for microbial monitoring of dry, sanitized surfaces and personnel in Clean Rooms and Isolators. The plate design allows aerobic, microaerophilic and anaerobic incubation.
- SDA + LTHTh - ICR+ selective (article number 146704, only available upon request): 90 mm lockable settle plates (transparent dishes); triple-bagged; gamma-irradiated; intended for microbial monitoring of air (passive and active) and fingerprints of personnel in Clean Rooms and Isolators. The plate design allows aerobic, microaerophilic and anaerobic incubation.

## Mode of Action

Sabouraud Dextrose Agar (SDA) is a complex medium for cultivation and isolation of yeasts and molds. The medium is supplemented with pyruvate to provide an efficient neutralization of hydrogen peroxide for use in Isolators. According to pharmacopoeia and ISO 18415, the neutralizers lecithin, polysorbate (Tween®) 80, histidine and sodium thiosulfate are suitable for neutralization of disinfectant residues containing the following active agents:

- Aldehydes
- Bis-biguanides
- Oxidizing compounds
- Parabens
- Phenolic compounds
- Quaternary ammonium compounds

The high concentration of dextrose in addition with the low pH promotes the growth, the formation of spores (conidia and sporangia) as well as the formation of pigments of yeasts and molds. In addition, the medium contains irradiation resistant antibiotics to inhibit the accompanying bacterial flora.

### Typical Composition

|                         |         |
|-------------------------|---------|
| Casein Peptone          | 5 g/l   |
| Meat Peptone            | 5 g/l   |
| Dextrose                | 40 g/l  |
| Polysorbate (Tween®) 80 | 5 ml/l  |
| Lecithin                | 0.7 g/l |
| Histidine               | 0.5 g/l |
| Sodium thiosulfate      | 0.3 g/l |
| Agar                    | 18 g/l  |
| Selective Supplements   |         |

The appearance of the medium is clear and yellowish. The pH value is in the range of 5.4-5.8. The medium can be adjusted and/or supplemented according to the performance criteria required.

### Application and Interpretation

The plates are introduced into Clean Rooms grade A or B by removing one bag in each material lock. For use in Isolators the inner bag has a hole in the sealing to hang up the bag during decontamination. Do not leave plates which are unprotected (unwrapped) in an Isolator during decontamination.

Each plate is provided with a label including a data matrix code for paperless plate identification. The code consists of a two-dimensional 20-digit serial number, which harbors the following information:

Digits 1-3: here code 706 (corresponds to article 146016); digits 4-9: lot number; digits 10-14: batch specific individual number; digits 15-20: expiry date (YY/MM/DD).

Please check each agar plate before using it on sterility and pay attention to aseptic handling to avoid false positive results.

The plates may be used for passive or active air monitoring as described in USP chapter <1116> or ISO 14698. For active air sampling please follow the guidance of the air sampler. Typically, 1000 liter of air are sampled for quantification of CFU. The exposure time of opened settle plates should be validated with respect to the environmental conditions of the sampling area such as flow rates, temperatures and relative humidity to preclude desiccation. Afterwards the plates are closed and transferred to an incubator. To protect the plates from secondary contamination during transport and incubation outside of the Clean Room zone, sterile transport bags (article number 146509) may be used.

Several recommendations are given by different guidelines for incubation: according to USP <1116> the plates used for environmental monitoring should be incubated between 20 and 35 °C for not less than 72 hours. According to the FDA Aseptic Guide the plates for determination of the total yeast and mold count should be incubated at 20 to 25 °C for 5 to 7 days. Individual incubation conditions can be chosen and should be validated at the application side.

Finally, the number of CFU per plate is examined.

Grown colonies are recommended to be identified.

### Storage and Shelf Life

The product can be used for sampling until the expiry date if stored upright, protected from light and properly sealed at +15 °C to +25 °C.

Condensation can be prevented by avoiding quick temperature shifts and mechanical stress.

The testing procedures as described on the CoA can be started up to the expiry date printed on the label.

### Disposal

Please mind the respective regulations for the disposal of used culture medium (e.g. autoclave for 20 min at 121 °C, disinfect, incinerate etc.).

### Quality Control

| Control Strains                                            | ATCC # | Inoculum CFU   | Incubation          | Expected Result Recovery in % |
|------------------------------------------------------------|--------|----------------|---------------------|-------------------------------|
| <i>Candida albicans</i><br>+ 100 µl Aerodesin 2000         | 10231  | 10-100         | 44-48 h at 20-25 °C | 50-200                        |
| <i>Aspergillus brasiliensis</i><br>+ 100 µl Aerodesin 2000 | 16404  | 10-100         | 70-74 h at 20-25 °C | 50-200                        |
| <i>Staphylococcus aureus</i><br>+ 100 µl Aerodesin 2000    | 6538   | 10,000-100,000 | 70-74 h at 20-25 °C | no growth                     |
| <i>Escherichia coli</i><br>+ 100 µl Aerodesin 2000         | 8739   | 10,000-100,000 | 70-74 h at 20-25 °C | no growth                     |

Please refer to the actual batch related Certificate of Analysis.

### Literature

EU GMP Medicinal Products for Human and Veterinary use (2008): Annex1 Manufacture of Sterile Medicinal Products.

European Pharmacopoeia 9.0 (2016): 2.6.12. Microbial examination of non-sterile products (total viable aerobic count).

Guidance for Industry (2004): Sterile Drug Products Produced by Aseptic Processing - Current Good Manufacturing Practice.

ISO 14698-1:2003: Clean Rooms and associated controlled environments - Biocontamination control - Part 1: General principles and methods.

ISO 18415 (2017 [E]): Cosmetics – Microbiology – Detection of specified and non-specified microorganisms

Japanese Pharmacopoeia 16<sup>th</sup> edition (2011): 4.05 Microbial Limit Test.

PDA Technical Report No. 13 (2014 Revised): Fundamentals of an Environmental Monitoring Program.



United States Pharmacopoeia 41 NF 36 (2018): <61> Microbiological Examination of Non-Sterile Products: Microbial Enumeration Tests; <1116> Microbiological Control and Monitoring of Aseptic Processing Environments.

## Ordering Information

| Product                                                  | Cat. No.     | Pack size          |
|----------------------------------------------------------|--------------|--------------------|
| Sabouraud Dextrose Agar + LTHTh - ICR selective          | 1.46016.0120 | 120 x 90 mm plates |
| Sabouraud Dextrose Contact Agar + LTHTh - ICR+ selective | 1.46538.0020 | 20 x 55 mm plates  |
| Sabouraud Dextrose Contact Agar + LTHTh - ICR+ selective | 1.46538.0200 | 200 x 55 mm plates |
| Sabouraud Dextrose Agar + LTHTh - ICR+ selective         | 1.46704.0020 | 20 x 90 mm plates  |
| Sabouraud Dextrose Agar + LTHTh - ICR+ selective         | 1.46704.0120 | 120 x 90 mm plates |
| Transport Bags, Sterile                                  | 1.46509.0125 | 25 x 5 bags        |

Merck KGaA  
64271 Darmstadt, Germany  
Fax: +49 (0) 61 51 / 72-60 80  
mibio@merckgroup.com

Find contact information for your country  
at: [www.merckmillipore.com/offices](http://www.merckmillipore.com/offices)

For Technical Service, please visit:  
[www.merckmillipore.com/techservice](http://www.merckmillipore.com/techservice)

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and liability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any right of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

Merck, Millipore, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. Detailed information on trademarks is available via publicly accessible resources.  
© 2019 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck operates as  
MilliporeSigma in the U.S. and Canada.

Lit. No. TN1268EN00

Page 4 of 4

The Merck logo, consisting of the word "MERCK" in a bold, blue, sans-serif font.