



Quality Control Ranges

MILLIPLEX[®] MAP Multi-Species Hormone Magnetic Bead Panel

Kit Catalog# MSHMAG-21K

QC Catalog# MSPHM-6021
Lot# MSHM-108 and MSHM-208

<u>Analyte</u>	<u>Level</u>	<u>Expected Range</u>	<u>Unit</u>
Cortisol	Control 1	2.1 – 4.4	ng/mL
	Control 2	23 – 47	ng/mL
Estradiol	Control 1	0.13 – 0.28	ng/mL
	Control 2	1.2 – 2.4	ng/mL
Progesterone	Control 1	0.92 – 1.9	ng/mL
	Control 2	11 – 24	ng/mL
T3	Control 1	1.69 – 3.5	ng/mL
	Control 2	17.6 – 37	ng/mL
T4	Control 1	1.5 – 3.1	ng/mL
	Control 2	23 – 48	ng/mL

MSPHM-6021: MSHM-108.208 (25 APRIL 2025)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • www.SigmaAldrich.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA